



Benannt durch Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
BS-MDR-099



Product Service

EU Technical Documentation Assessment Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapter II
(Implantable Class IIb Devices and Class III Devices)

No. G70 078611 0183 Rev. 00

Manufacturer:

Cochlear Limited

1 University Avenue
Macquarie University NSW 2109
AUSTRALIA

SRN Manufacturer - AU-MF-000009890

Authorized Representative:

Cochlear Deutschland GmbH & Co. KG
Mailänder Straße 4 a, 30539 Hannover, GERMANY

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has drawn up and presented a Technical Documentation according to Annex II and III of the Regulation (EU) 2017/745 on medical devices. Details on devices covered by the Technical Documentation are described on the following page(s)

The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The technical documentation assessment included an assessment of the clinical evaluation assessment.

The conformity assessment has been carried out according to Annex IX chapter II of this regulation with a positive result.

Changes to the approved device, where such changes could affect the safety and performance of the device or the conditions prescribed for use of the device, shall require approval from the notified body TÜV SÜD Product Service GmbH.

In order to place the devices on the market with CE-marking, an EU Quality Management System Certificate pursuant to Annex IX chapters I and III is necessary in addition to this EU Technical Documentation Assessment Certificate. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see:

www.tuvsud.com/ps-cert?q=cert:G70_078611_0183_Rev.00

Report No.: 713282650

Valid from: 2023-11-09

Valid until: 2028-11-08

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2023-11-09



Benannt durch Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 BS-MDR-099



Product Service

EU Technical Documentation Assessment Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapter II
 (Implantable Class IIb Devices and Class III Devices)

No. G70 078611 0183 Rev. 00

Classification: Class III
Device Group: J0380 - AUDITORY ACTIVE-IMPLANTABLE DEVICES - ACCESSORIES
Basic UDI-DI: 9321502CP1150SP58
Intended Purpose: The sound processor is intended to be used in combination with other devices as part of a hearing implant system to provide hearing sensation. The sound processor converts sounds into electrical signals, which it sends to an implant. The sound processor also provides power to the implant.
Device(s): Kanso® 2 Sound Processor, model CP1150. Available in the following variants:
 Kanso® 2 Sound Processor, CP1150 - Sandy Blonde
 Kanso® 2 Sound Processor, CP1150 - Silver
 Kanso® 2 Sound Processor, CP1150 - Chocolate Brown
 Kanso® 2 Sound Processor, CP1150 - Slate Grey
 Kanso® 2 Sound Processor, CP1150 - Black

The validity of this certificate depends on conditions and/or is limited to the following: -none-

Revision History:

Rev.	Dated	Report	Description
00	2023-11-09	713282650	Initial issuance